




THE TOP IN QUALITY  
for all areas of healthcare

# MULTI SYRINGE PUMP STAND POLE CLAMP



TOP MULTIPLE SYRINGE PUMP STAND  
(Multi-pole clamp)



TOP POLE CLAMP

## TOP MULTIPLE SYRINGE PUMP STAND

- Stacking / Docking station for 4 units of TOP Syringe Pumps
- One-touch function.
- Power supply is by one AC Power Cable only.

## POLE CLAMP

- One-touch function.

| Product                         | Identifier | Target model                          | Package   |
|---------------------------------|------------|---------------------------------------|-----------|
| TOP MULTIPLE SYRINGE PUMP STAND | ST55-4     | TOP Syringe Pump<br>(TOP-55xx series) | 1unit/box |
| TOP POLE CLAMP                  | TYPE E2    |                                       | 1pc./box  |

ASAL  
ORIGINAL

PIHAK BERKUASA  
PERANTI PERUBATAN



MEDICAL DEVICE  
AUTHORITY

**PIHAK BERKUASA PERANTI PERUBATAN**  
MEDICAL DEVICE AUTHORITY  
**AKTA PERANTI PERUBATAN 2012 (AKTA 737)**  
MEDICAL DEVICE ACT 2012 (ACT 737)  
**SIJIL PENDAFTARAN PERANTI PERUBATAN**  
MEDICAL DEVICE REGISTRATION CERTIFICATE  
**Seksyen 5(1) Akta 737**  
Section 5(1) of Act 737

No. Pendaftaran: **GMD73175393318A**  
Registration No.:

Tarikh Sah Laku Pendaftaran:  
Registration Validity Date:

**08/01/2023 -**  
**07/01/2028**

Sijil ini adalah dengan ini diberi kepada:  
*This certificate is hereby issued to:*

**MEDITOP CORPORATION (MALAYSIA) SDN. BHD.**

yang beralamat di:  
*which is located at:*

**NO. 3, PERSIARAN USAHAWAN, TAMAN IKS,  
SEKSYEN 9, ,  
43650 BANDAR BARU BANGI  
SELANGOR DARUL EHSAN**

bagi mengesahkan peranti perubatan seperti yang dinyatakan dalam Lampiran 1 adalah berdaftar di bawah Seksyen 5(1) Akta 737.  
*to confirm that the medical device as detailed out in Attachment 1 is registered under Section 5(1) of Act 737.*

Pendaftaran ini diberikan tertakluk kepada peruntukan-peruntukan di bawah Akta 737 dan peraturan-peraturan yang dibuat dibawahnya serta syarat-syarat seperti di Lampiran 2.  
*This registration is granted subject to the provisions under Act 737 and its subsidiary legislations and the conditions as in Attachment 2.*



**AHMAD SHARIFF BIN HAMBALI**  
**KETUA EKSEKUTIF**  
CHIEF EXECUTIVE  
**PIHAK BERKUASA PERANTI PERUBATAN**  
MEDICAL DEVICE AUTHORITY



No. Pendaftaran: **GMD73175393318A**  
Registration No.:

Butir-butir peranti perubatan yang didaftarkan  
Particulars of the registered medical device

Nama Peranti Perubatan **TOP MULTIPLE SYRINGE PUMP STAND**  
Medical Device Name

|                   |                |                                           |               |
|-------------------|----------------|-------------------------------------------|---------------|
| Kelas<br>Class    | <b>CLASS A</b> | Jenama<br>Brand                           | <b>TOP</b>    |
| Kelompok<br>Group | <b>SINGLE</b>  | Produk Identifikasi<br>Product Identifier | <b>ST55-4</b> |

|                                                                    |                                                                                        |
|--------------------------------------------------------------------|----------------------------------------------------------------------------------------|
| Nama dan alamat<br>pembuat:<br>Name and address of<br>manufacturer | <b>TOP CORPORATION<br/>19-10 SENJU NAKAI-CHO, ADACHI-KU, TOKYO 120-0035,<br/>JAPAN</b> |
|--------------------------------------------------------------------|----------------------------------------------------------------------------------------|

**Medical Device**

AUTHORITY

MALAYSIA

## SYARAT – SYARAT PENDAFTARAN PERANTI PERUBATAN MEDICAL DEVICE REGISTRATION CONDITIONS

### 1.0 SYARAT AM GENERAL CONDITIONS

- 1.1 Syarat-syarat pendaftaran peranti perubatan ini dibuat adalah berdasarkan kepada Seksyen 7 (1), Akta Peranti Perubatan 2012 (Akta 737). Kelulusan ini diberi berdasarkan maklumat-maklumat yang telah diterima.  
*Medical device registration conditions are prescribed in accordance to Section 7(1) of Medical Device Act (Act 737). Approval is granted based on information received.*
- 1.2 Establismen hendaklah mematuhi segala arahan yang dikeluarkan oleh Pihak Berkuasa dari semasa ke semasa.  
*Establishment must comply with all instructions issued by the Authority from time to time.*
- 1.3 Pihak Berkuasa berhak untuk meminda syarat-syarat pendaftaran dari semasa ke semasa.  
*The Authority reserves the rights to amend the registration conditions from time to time.*
- 1.4 Pihak Berkuasa berhak menjalankan pemeriksaan ke atas establismen pada bila-bila masa tanpa dimaklumkan terlebih dahulu.  
*The Authority reserves the right to conduct inspection at any time without prior notice.*
- 1.5 Pihak Berkuasa boleh membatalkan Pendaftaran Peranti Perubatan atau mengambil tindakan undang-undang sekiranya Establismen gagal mematuhi mana-mana syarat Pendaftaran Peranti Perubatan.  
*The Authority may cancel the Medical Device Registration or take legal action if the Establishment fails to comply with any medical device registration conditions.*
- 1.6 Sijil Pendaftaran Peranti Perubatan yang dikeluarkan oleh Pihak Berkuasa tidak boleh dipindah milik.  
*Medical Device Registration Certificate issued by the Authority shall not be transferable or assignable.*
- 1.7 Sijil Pendaftaran Peranti Perubatan hendaklah dikemukakan sekiranya diminta oleh mana-mana pegawai yang diberi kuasa.  
*Medical Device Registration Certificate must be presented upon request by any authorized officer.*
- 1.8 Establismen tidak boleh membenarkan Sijil Pendaftaran Peranti Perubatan disalahgunakan oleh individu/syarikat lain dalam apa-apa cara.  
*Establishment shall not permit the Medical Device Registration Certificate to be abused in any way by any individual / another party.*
- 1.9 Tempoh sahlaku Sijil Pendaftaran Peranti Perubatan adalah lima (5) tahun dari tarikh pendaftaran melainkan jika pendaftaran itu dibatalkan oleh Pihak Berkuasa sebelum habis tempohnya.  
*The validity of the Medical Device Registration Certificate is five (5) years from the date of registration unless the registration is cancelled by the Authority before its expiry.*
- 1.10 Sijil ini tidak mengecualikan mana-mana keperluan perundangan lain yang terpakai untuk sesuatu peranti perubatan (contoh: Peranti Perubatan yang mengandungi racun berjadual tertakluk kepada Akta Racun 1952; peranti perubatan menggunakan sinaran mengion adalah tertakluk kepada Akta Perlesenan Tenaga Atom 1984.)  
*This certificate does not exempt any other regulatory requirements applicable to the medical device (for examples: Medical Device containing scheduled poison is subjected to the Poisons Act 1952; medical devices using ionizing radiation is subjected to the Atomic Energy Licensing Act 1984 ) .*

**1.11** Establismen hendaklah melaporkan insiden melibatkan peranti perubatan yang didaftarkan kepada Pihak Berkuasa seperti tertakluk di bawah Seksyen 40 Akta 737.  
*Establishment shall report any incidents involving registered medical device to the Authority as prescribed in Section 40 of Act 737.*

**1.12** Peranti perubatan yang diniatkan bagi kegunaan professional hanya boleh dibekalkan untuk kegunaan professional perubatan sahaja dan tidak boleh diletakkan dipasaran bagi kegunaan orang awam.  
*Medical device intended for professional use may only be supplied for use by medical professionals only and shall not be placed in the market for general public.*

## **2.0 PINDAAN PENDAFTARAN PERANTI PERUBATAN AMENDMENT OF MEDICAL DEVICE REGISTRATION**

**2.1** Sebarang pindaan kepada maklumat yang berkaitan peranti perubatan yang berdaftar hendaklah dimaklumkan kepada Pihak Berkuasa secara rasmi mengikut garis panduan yang ditetapkan oleh Pihak Berkuasa. Pihak Berkuasa berhak memberikan kelulusan atau menolak permohonan pindaan tersebut.  
*Any amendments to the information concerning registered medical device shall be notified to the Authority in accordance to the guidelines set by the Authority. The Authority reserves the right to grant approval or reject the application for such amendments.*

## **3.0 PEMBATALAN SIJIL PENDAFTARAN PERANTI PERUBATAN CANCELLATION OF MEDICAL DEVICE REGISTRATION CERTIFICATE**

**3.1** Sijil Pendaftaran Peranti Perubatan boleh dibatalkan seperti yang dinyatakan dalam Seksyen 9, Akta 737.  
*Medical Device Registration certificate may be cancelled as prescribed in Section 9 of Act 737.*

**3.2** Mana-mana peranti perubatan yang dibatalkan Sijil Pendaftarannya, tidak boleh diimport, dieksport atau diletakkan dalam pasaran.  
*Any Medical Device which the registration certificate has been cancelled shall not be imported, exported or placed in the market.*

**3.3** Sijil Pendaftaran Peranti Perubatan boleh dibatalkan jika Wakil Diberi Kuasa ditamatkan lantikan oleh pembuat.  
*Medical Device Registration Certificate may be cancelled if Authorized Representative appointment is terminated by the manufacturer.*

## **4.0 HAK PIHAK BERKUASA THE AUTHORITY OWNERSHIP**

**4.1** Sijil Pendaftaran Peranti Perubatan yang dikeluarkan fizikal atau maya adalah **Hak Milik Pihak Berkuasa**.  
*The Authority retains the ownership of every Medical Device Registration Certificate issued by any means.*

**4.2** Sekiranya berlaku kehilangan atau kerosakan Sijil Pendaftaran Peranti Perubatan, hendaklah dimaklumkan kepada Pihak Berkuasa dan setiap penggantian sijil akan dikenakan caj perkhidmatan.  
*Any loss or damage to the Medical Device Registration Certificate shall be notified to the Authority and every replacement of certificate shall be liable with service charge rendered.*

## **5.0 TUGAS DAN TANGGUNGJAWAB ROLES AND RESPONSIBILITIES**

**5.1** Establismen hendaklah mematuhi Akta 737, peraturan-peraturan di bawah Akta dan syarat-syarat Pendaftaran Peranti Perubatan.  
*Establishment shall comply with Act 737, its subsidiary regulations and registration Condition*